



## Application

Ideal for calibrating thermocouples, RTD, indicators, controllers, transmitters, receivers and mA, mV, V, ohms measuring instruments

## Features

- Graphical Display
- Microcontroller Based
- Loop Power Supply for Transmitters
- Simulation and Measurement of RTD, thermocouple, mA, mV or V,  $\Omega$
- Sink facility in mA mode
- Internal Cold Junction Compensation
- Light weight, handy & portable
- Feather Touch Key Pad



## Specification - Measurement Mode

Parameter	Range	Resolution	Accuracy <sup>1</sup>
mA	0 to 25	0.001 (mA)	$\pm 0.025\%$ of Reading + 10 $\mu$ A
mV	-10 to 200	0.001 (mV)	$\pm 0.025\%$ of Reading + 10 $\mu$ V
V	0 to 10	0.001 (V)	$\pm 0.025\%$ of Reading + 3 mV
Ohms <sup>2</sup>	0 to 400	0.01 (Ohm)	$\pm 0.025\%$ of Reading + 0.1 Ohm
RTD (Pt 100) <sup>3</sup>	-200 to 850	0.1 ( $^{\circ}$ C)	$\pm 0.025\%$ of Reading + 0.3 $^{\circ}$ C

### Thermocouple<sup>3</sup> $^{\circ}$ C

J	Range	Resolution	Accuracy
J	-200 to 1200	0.1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 1 $^{\circ}$ C
K	-200 to 1370	0.1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 1 $^{\circ}$ C
T	-200 to 400	0.1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 1 $^{\circ}$ C
E	-200 to 1000	0.1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 1 $^{\circ}$ C
N	-200 to 1300	0.1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 1 $^{\circ}$ C
R	300 to 1750	1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 2 $^{\circ}$ C
S	300 to 1750	1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 2 $^{\circ}$ C

Note 1: Accuracy specifications include effect of non-linearity, hysteresis and repeatability at 23 $\pm$ 5 $^{\circ}$ C

Note 2: Measurement by 3 wire at an excitation current of 1mA

Note 3: Thermocouple measurement accuracy does not include CJC sensor accuracy (CJC error  $\pm 1.0^{\circ}$ C).

Note 4: Maximum load 1000 ohms

Note 5: Source excitation current 0.1 to 2mA

## Specification - Source Mode

Parameter	Range	Resolution	Accuracy <sup>1</sup>
mA <sup>4</sup>	0 to 25	0.001 (mA)	$\pm 0.025\%$ of Reading + 10 $\mu$ A
mV	-10 to 200	0.01 (mV)	$\pm 0.025\%$ of Reading + 10 $\mu$ V
V	0 to 10	0.001 (V)	$\pm 0.025\%$ of Reading + 3 mV
Ohms <sup>5</sup>	0 to 400	0.01 (Ohm)	$\pm 0.025\%$ of Reading + 0.1 Ohm
RTD (Pt 100) <sup>5</sup>	-200 to 850	0.1 ( $^{\circ}$ C)	$\pm 0.025\%$ of Reading + 0.3 $^{\circ}$ C

### Thermocouple<sup>3</sup> $^{\circ}$ C

J	Range	Resolution	Accuracy
J	-200 to 1200	0.2 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 2 $^{\circ}$ C
K	-200 to 1370	0.2 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 2 $^{\circ}$ C
T	-200 to 400	0.2 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 2 $^{\circ}$ C
E	-200 to 1000	0.2 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 2 $^{\circ}$ C
N	-200 to 1300	0.2 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 2 $^{\circ}$ C
R	300 to 1750	1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 3 $^{\circ}$ C
S	300 to 1750	1 $^{\circ}$ C	$\pm 0.025\%$ of Reading + 3 $^{\circ}$ C

### Current Terminal Input

Protection	: Fuse: 100mA / 250V
Memory Function	: 9 set of values for each parameter in generation
Backlight	: LED Backlight
Power Supply	: Li-ion Rechargeable Battery
Low Battery Indicator	: Yes
Battery Backup	: 18 hours @ no load condition 11 hours 30 mins. @ 20mA with 850 ohm load
Loop Power Supply	: 24V $\pm 10\%$
Operating Temperature	: 10 to 50 $^{\circ}$ C
Instrument Casing	: ABS Plastic
Instrument Dimensions	: 250 mm x 140 mm x 45 mm (LxWxD)
Instrument Weight	: 500gms

### Standard Delivery

1. Instrument : 1 no.
2. Carrying Pouch : 1 no.
3. Test Lead (2 wire) : 2 nos.
4. Test Lead (3 wire) : 1 no.
5. Charger : 1 no.
6. User Manual : 1 no.
7. Traceable Calibration Certificate

### Optional

Accredited (NABL) Calibration Certificate as per ISO/IEC 17025 Standard

Manufacturer of  
**TEST BENCHES  
CALIBRATION INSTRUMENTS**

**ISO 9001  
Certified Company**

[www.rndinstruments.com](http://www.rndinstruments.com)

### Contact

5, Natesan Nagar, Alapakkam,  
Chennai 600116. Tamilnadu, INDIA.  
Tel.: +91-44-4283 9039 / 4284 1119  
Mobile Nos.: +91-7358393981 / 7358393984  
email: sales@rndinstruments.com  
service@rndinstruments.com

**ACCREDITED CALIBRATION LABORATORY**

As per ISO / IEC 17025

**THERMAL  
MECHANICAL  
ELECTROTECHNICAL**

[www.calibrationlab.in](http://www.calibrationlab.in)